

ACCEPTED MANUSCRIPT • OPEN ACCESS

Alanine dosimetry in strong magnetic fields: use as a transfer standard in MRI-guided radiotherapy

To cite this article before publication: Ilias Billas *et al* 2020 *Phys. Med. Biol.* in press <https://doi.org/10.1088/1361-6560/ab8148>

Manuscript version: Accepted Manuscript

Accepted Manuscript is “the version of the article accepted for publication including all changes made as a result of the peer review process, and which may also include the addition to the article by IOP Publishing of a header, an article ID, a cover sheet and/or an ‘Accepted Manuscript’ watermark, but excluding any other editing, typesetting or other changes made by IOP Publishing and/or its licensors”

This Accepted Manuscript is © 2020 Institute of Physics and Engineering in Medicine.

As the Version of Record of this article is going to be / has been published on a gold open access basis under a CC BY 3.0 licence, this Accepted Manuscript is available for reuse under a CC BY 3.0 licence immediately.

Everyone is permitted to use all or part of the original content in this article, provided that they adhere to all the terms of the licence <https://creativecommons.org/licenses/by/3.0>

Although reasonable endeavours have been taken to obtain all necessary permissions from third parties to include their copyrighted content within this article, their full citation and copyright line may not be present in this Accepted Manuscript version. Before using any content from this article, please refer to the Version of Record on IOPscience once published for full citation and copyright details, as permissions may be required. All third party content is fully copyright protected and is not published on a gold open access basis under a CC BY licence, unless that is specifically stated in the figure caption in the Version of Record.

View the [article online](#) for updates and enhancements.

Alanine dosimetry in strong magnetic fields: use as a transfer standard in MRI-guided radiotherapy

Ilias Billas^{1,3}, Hugo Bouchard², Uwe Oelfke³, David Shipley¹, Clare Gouldstone¹ and Simon Duane¹

¹*National Physical Laboratory, Teddington, UK*

²*Université de Montréal, Département de physique, Montréal, Canada*

³*Joint Department of Physics, The Institute of Cancer Research and The Royal Marsden NHS
Foundation Trust, London, UK*

Keywords: MRI-guided radiotherapy, magnetic field, reference dosimetry, Alanine

1 ABSTRACT

Reference dosimetry in the presence of a strong magnetic field is challenging. Ionisation chambers have shown to be strongly affected by magnetic fields. There is a need for robust and stable detectors in MRI-guided radiotherapy (MRIgRT). This study investigates the behaviour of the alanine dosimeter in magnetic fields and assess its suitability to act as a reference detector in MRIgRT.

Alanine pellets were loaded in a waterproof holder, placed in an electromagnet and irradiated by ⁶⁰Co and 6 MV and 8 MV linac beams over a range of magnetic flux densities. Monte Carlo simulations were performed to calculate the absorbed dose, to water and to alanine, with and without magnetic fields. Combining measurements with simulations, the effect of magnetic fields on alanine response was quantified and a correction factor for the presence of magnetic fields on alanine was determined.

This study finds that the response of alanine to ionising radiation is modified when the irradiation is in the presence of a magnetic field. The effect is energy independent and may increase the alanine/EPR signal by 0.2% at 0.35 T and 0.7% at 1.5 T. In alanine dosimetry for MRIgRT, this effect, if left uncorrected, would lead to an overestimate of dose. Accordingly, a correction factor, $k_{QB,Q}$, is defined. Values are obtained for this correction as a function of magnetic flux density, with a standard uncertainty which depends on the magnetic field and is 0.6% or less.

The strong magnetic field has a measurable effect on alanine dosimetry. For alanine which is used to measure absorbed dose to water in a strong magnetic field, but which has been calibrated in the absence of a magnetic field, a small correction to the reported dose is required. With the inclusion of this correction, alanine/EPR is a suitable reference dosimeter for measurements in MRIgRT.

2 INTRODUCTION

MRI-guided radiotherapy (MRIgRT), a state-of-the-art cancer treatment, is delivered by an MRI-linac which combines a linear accelerator (linac) with a Magnetic Resonance Imaging (MRI) scanner. An MRI-linac provides real-time images during a patient's treatment and greatly enhanced soft-tissue image contrast, while completely avoiding the radiation dose associated with treatment integrated imaging X-ray systems. MRIgRT is expected to facilitate real-time adaptive radiotherapy (RT) planned based on high contrast anatomical images, visualising anatomical changes of the patient, to explore the possibilities of an advanced personalised radiotherapy.

A novel but known feature of this technology is that the external magnetic field affects the response of dosimeters and the dose distribution in medium. The lack of a standard for reference dosimetry of RT beams in magnetic field, motivates our investigation.

Reference dosimetry for conventional radiotherapy is usually performed with ionisation chambers. In practice, secondary chambers are calibrated against a primary standard for absorbed dose to water (based on ionometry, water calorimetry, graphite calorimetry, etc.) and then used to calibrate the output of radiotherapy machines. In the presence of a magnetic field, the charged particles traversing the air-cavity of an ionisation chamber experience the additional Lorentz force induced by the magnetic field, which strongly modifies the chambers dose response and makes it challenging to perform beam output measurements in MRI-linacs. Recent works have shown that the uncertainty of ionisation chamber-based dosimetry is significantly increased by the presence of strong magnetic fields. Measurements and Monte Carlo (MC) calculations have previously been made to characterise the response of different type of ionisation chambers (Meijsing *et al.*, 2009, Smit *et al.*, 2013, O'Brien *et al.*, 2016, Reynolds *et al.*, 2013, Spindeldreier *et al.*, 2017, Pojtinger *et al.*, 2018). These works investigated the optimal chamber orientation with respect to magnetic field and radiation beam, parallel (||) or perpendicular (\perp), as well as the magnetic field correction factor at different flux densities. Considering the change of a Farmer-type chamber (mostly used as a reference detector) response in magnetic flux densities ranging from 0.35 T to 1.5 T relative to its response at 0 T, these studies found that when the long axis of the ionisation chamber is:

- \perp to the magnetic field and the radiation beam is \perp to the magnetic field the changes range from 4% - 11.3%.
- || to the magnetic field and the radiation beam is \perp to the magnetic field the change is $\leq 1\%$.
- || or \perp to the magnetic field and the radiation beam is || to the magnetic field the change increases near to 2%.

In the presence of the strong magnetic fields the radius of the circular motion of a charged particle vary with the mass density. The higher the mass density the higher the radius of the particle trajectory. Therefore, although the effect of the magnetic field is more pronounced in air cavities, it is also affecting the distribution of dose in water (Raaijmakers *et al.*, 2004, Raaijmakers *et al.*, 2007, Raaijmakers *et al.*, 2005, O'Brien *et al.*, 2016). A study by O'Brien *et al.* (2016) characterised depth-dose curves with an Elekta Unity MR-linac beam quality and showed that the presence of the magnetic field changes the absorbed dose at a reference depth by -0.5%. This effect needs to be taken into account and corrected for or it can be a part of the quality correction factor for the presence of magnetic field, $k_{Q_B,Q}$, which has been discussed by O'Brien *et al.* (2016) and van Asselen *et al.* (2018) and is further described on section 3.2 of this paper. Both mentioned studies have used the analogy of radiation beam quality correction factor, which is defined as the ratio between the calibration coefficient in standard conditions and the calibration coefficient in user conditions, as stated on TRS 398 (Andreo *et al.*, 2000). In section 3.2, $k_{Q_B,Q}$ is expressed as a product of the effect of the magnetic field on dose to water and on detector response. For different Farmer-type chambers, $k_{Q_B,Q}$ ranges from 0.992 to 1.006 when the long axis of the chamber is parallel to magnetic field and 0.953 to 0.976 when the long axis is directed perpendicular to magnetic field.

Among other effects of ionisation chambers in magnetic fields, which increase the uncertainty on dose measurement, are the air gaps around the cavity (Hackett *et al.*, 2016, Agnew *et al.*, 2017). This will restrict the use of ionisation chambers in water phantoms to avoid a significant signal variation, i.e. 3.8% for a Farmer-type chamber (Agnew *et al.*, 2017). It has been shown by Malkov and Rogers (2017), Spindeldreier *et al.* (2017) and Pojtinger *et al.* (2019) that dead volumes (or the true sensitive volume) should be taken into account when performing MC simulations. Small changes in the collecting volumes can lead to a deviation of 1.4% between the simulated and the experimentally determined dose response in magnetic fields (Pojtinger *et al.*, 2019), on Farmer-type chambers. This,

together with the intra-type chamber variation (i.e. non-radial symmetry), will result in significant disagreement with experimental data and will increase the uncertainty on tabulated MC calculated quality correction factors for the presence of magnetic field for specific chambers.

Alanine is an α -amino acid with the chemical formula: $\text{CH}_3\text{-CH}(\text{NH}_2)\text{-COOH}$. When an alanine molecule is irradiated, it produces a stable free radical, $\text{CH}_3\text{-C}\cdot\text{H-COOH}$. The concentration of the stable free radicals can be measured as a signal (a dosimetric signal) using Electron Paramagnetic Resonance (EPR) spectroscopy and is proportional to the absorbed dose. A detailed description of alanine as a dosimeter has been given in the literature (McEwen and Ross, 2009, Malinen, 2014) and some primary points have been emphasised by McEwen M. *et al.* (2015).

Alanine dosimeters provide an alternative to standard dosimetry systems, such as ionisation chambers. McEwen M. *et al.* (2015) has pinpointed the advantages of alanine dosimetry as a reference detector against an ionisation chamber in conventional radiotherapy. The National Physical Laboratory (NPL) provides an alanine reference dosimetry service (Sharpe *et al.*, 1996, Sharpe and Sephton, 2000, Sharpe and Sephton, 2006), which can be used to measure therapy level absorbed dose in reference and non-reference conditions. Alanine has been used as a transfer standard for comparison of the absorbed dose measured from national standards laboratories in a Co-60 beam (Burns *et al.*, 2011) and in megavoltage electron beams (McEwen M. *et al.*, 2015) with a precision of 0.1%. The use of alanine in dosimetry audits for megavoltage photons and electrons under standard conditions has been proven by Thomas *et al.* (2003) to be an essential detector to ensure national consistency in radiotherapy dosimetry. In non-standard conditions, such as small field beams (Massillon *et al.*, 2013, Ramirez *et al.*, 2011, Dimitriadis *et al.*, 2017) and flattening filter free beams (Budgell *et al.*, 2016) alanine has been used for absolute dosimetry measurements transferring traceability to a primary standard.

Currently, the only primary standard that can provide traceability for reference dosimetry in magnetic field is the VSL's water calorimeter (de Prez *et al.*, 2019). To our knowledge, this is the only MRI-safe primary standard, which can also fit in the bore of the MRI scanner, that has performed direct measurements of absorbed dose to water in an Elekta Unity MR-linac and cross calibrated ionisation chambers to act as transfer standards. Nevertheless, as already explained above, the strong magnetic field modifies the response of air-filled ionisation chambers, and the correction $k_{Q_B,Q}$ for this effect may significantly increase the uncertainty of reference dosimetry. On the other hand, alanine/EPR offers an alternative to ionisation chamber-based reference dosimetry, in which the magnitude of the electron return effect (ERE) may be reduced compared to the effect in an air-filled ionisation chamber.

Taking into account a) the disadvantages of the ionisation chamber response in magnetic field, b) the need to find a traceable dosimeter which will allow for MR safety, robustness, stability and dose response and c) the wide use of alanine as a reference detector for standard and non-standard radiation beams, the aim of this work is to investigate the performance of alanine in the presence of strong magnetic field and to explore its suitability to act as a reference class detector for MRIgRT systems.

The use of alanine as a transfer standard implies that it is used in the same conditions that it has been calibrated (primary condition). Where these conditions do not match the measurement conditions (i.e. alanine is irradiated in a different beam quality) a correction to account for changes in the response relative to those under calibration conditions need to be applied. For instance, the alanine/EPR signal depends on the energy spectrum of the radiation beam. Corrections for this effect have been well quantified in the literature (Bergstrand *et al.*, 2005, Zeng *et al.*, 2004, Sharpe, 2003,

Bergstrand *et al.*, 2003, Anton *et al.*, 2013). When alanine is used to measure dose in the presence of a strong magnetic field (such as in MRI-linac systems), and an alanine calibration is available only in the absence of the magnetic field, then a correction factor that would account for the effect of the magnetic field needs to be applied. This factor will correct for both the effect of the magnetic field on dose to water and on alanine response. In the present study, we perform measurements and MC simulations to quantify the effect of the magnetic field on alanine response (described as alanine intrinsic sensitivity) and generate magnetic field correction factors for alanine. A statistical analysis, including an investigation of the effect of the air gaps associated with the alanine holder, is also presented.

3 MATERIALS AND METHODS

3.1 ALANINE DOSIMETRY

Alanine pellets comprise of 90% L- α -alanine and 10% high melting point paraffin wax (m.p. 98 °C), by weight (Sharpe *et al.*, 1996). Alanine is near water-equivalent, but not waterproof, with a density of about 1.2 g/cm³ and it has a linear response to dose. It also has good reproducibility (0.3% variation between individual dosimeter pellets) and a small energy dependence, less than 1% relative to a ⁶⁰Co beam (Bergstrand *et al.*, 2005, Zeng *et al.*, 2004, Sharpe, 2003, Bergstrand *et al.*, 2003, Anton *et al.*, 2013). EPR spectroscopy is used to detect and quantify the number of the stable free radicals that are created by the ionising radiation. The alanine/EPR signal is the corrected alanine peak-to-peak intensity per unit dosimeter mass, expressed in Gy to water, using a calibration curve derived from reading out alanine irradiated under reference conditions in ⁶⁰Co. The corrections include effects such as temperature during irradiation and spectrometer sensitivity. If alanine is irradiated in some other conditions, then a correction may be needed to account for any change in the intrinsic sensitivity of alanine, which may be defined as the alanine/EPR signal per unit absorbed dose to alanine.

3.2 ALANINE REFERENCE DOSIMETRY IN A MAGNETIC FIELD

The reference measurement of absorbed dose to water, in the presence of a magnetic field B , D_{w,Q_B} , using alanine, is given by:

$$D_{w,Q_B} = M_{al,Q_B} \cdot N_{D,w,Q_B} \quad (1)$$

where, M_{al,Q_B} and N_{D,w,Q_B} are the alanine/EPR signal and the calibration coefficient, respectively, of the alanine detector in the presence of magnetic field. The magnetic field affects the absorbed dose to water at the measurement point, $D_{w,Q}$, the absorbed dose to the alanine pellet, $D_{al,Q}$, and may also affect the intrinsic sensitivity of the alanine, $R_{al,Q}$, which is defined by:

$$R_{al,Q} = \frac{M_{al,Q}}{D_{al,Q}} \quad (2)$$

Note that this relation is analogous to the definition of detector dose response (or sensitivity) proposed by Bouchard and Seuntjens (2013).

If the alanine calibration is only available for a beam quality Q in the absence of any magnetic field, $N_{D,w,Q}$, then a correction factor for the effect of the magnetic field is needed:

$$D_{w,Q_B} = M_{al,Q_B} \cdot N_{D,w,Q} \cdot k_{Q_B,Q} \quad (3)$$

This correction factor is defined, by analogy with the usual quality dependent correction factor, k_{Q,Q_0} (Andreo *et al.*, 2000), by:

$$k_{Q_B,Q} = \frac{N_{D,w,Q_B}}{N_{D,w,Q}} = \frac{D_{w,Q_B}/M_{al,Q_B}}{D_{w,Q}/M_{al,Q}} = \frac{D_{w,Q_B}}{D_{w,Q}} \cdot \frac{M_{al,Q}}{M_{al,Q_B}} \quad (4)$$

The absorbed dose to water, with magnetic field, D_{w,Q_B} , and without magnetic field, $D_{w,Q}$, was calculated by using MC simulations. The alanine/EPR signal at a quality Q , $M_{al,Q}$, can be derived from equation (2) as the product of the absorbed dose to alanine at a quality Q , $D_{al,Q}$, and the alanine intrinsic sensitivity at the same quality Q , $R_{al,Q}$. The effect of the magnetic field on the alanine intrinsic sensitivity may be represented by the ratio, $F_{Q_B,Q}$, and is defined as relative intrinsic sensitivity:

$$F_{Q_B,Q} = \frac{R_{al,Q_B}}{R_{al,Q}} \quad (5)$$

This was determined by combining measurements of the alanine/EPR signal with the absorbed dose to the alanine pellet, determined by MC simulation of the dosimeter setup:

$$F_{Q_B,Q} = \frac{M_{al,Q_B}}{M_{al,Q}} \cdot \frac{D_{al,Q}}{D_{al,Q_B}} = \frac{D_{w,Q_B}/D_{w,Q} \cdot D_{al,Q}/D_{al,Q_B}}{k_{Q_B,Q}} \quad (6)$$

3.3 EXPERIMENTAL SETUP

Alanine pellets, of approximately 2.3 mm height and 5 mm diameter, were placed in a waterproof holder of Polyether Ether Ketone (PEEK) material shaped like a Farmer-type ionisation chamber, in an electromagnet (GMW 3474-140) and irradiated by either a ^{60}Co source or conventional 6 MV and 8 MV Elekta Synergy linac beams. A disassembled view of the NPL alanine holder is shown in Figure 1. The external dimensions of this holder are designed to match a PTW/30013 Farmer chamber. The holder has internal dimensions of 5.2 mm diameter (tolerance of 0.2 mm to allow the loading and unloading of the pellets) and a length of 20.5 mm. The measurement reference point of the PTW/30013 Farmer-type chamber is very close to the centre of the third pellet from the tip of the holder. Thus, the first five pellets from the tip of the holder were used for analysis.



Figure 1: NPL alanine holder.

In ^{60}Co the alanine dosimeters were irradiated in an in-house poly-methyl methacrylate (PMMA) phantom, 2 cm x 5 cm x 26 cm, with a drilled insert to enable the alanine holder to fit in. Any air gap between the insert and the holder was carefully filled with water to avoid the ERE. The phantom was designed to fit in the 5 cm gap between the two poles of the magnet, which provide a maximum magnetic flux density of 2 T. The distance between the ^{60}Co source to the centre of the magnetic poles

(measurement point) was 162 cm and alanine pellets were irradiated at a depth of 1 cm at five magnetic flux densities (0 T, 0.5 T, 1 T, 1.5 T and 2 T).

In the Elekta Synergy linac setup, a water phantom, 7 cm x 18 cm x 19 cm, was designed to place the alanine holder between the 7 cm gap of the magnetic poles (Figure 2), which provide a maximum magnetic flux density of 1.6 T. An insert frame-holder was built to set up the alanine holder in the water tank and between the two poles of the magnet. The surface to source distance (SSD) was 305 cm, with alanine irradiated at a water equivalent depth of 5 cm and magnetic flux densities of 0 T, 0.35 T, 0.5 T, 1 T and 1.5 T.




Figure 2: Front schematic view (cross-section) of the experimental setup. PMMA or water phantom fits between the two magnetic poles with a gap of either 5 cm or 7 cm that gives a maximum magnetic flux density of either 2 T or 1.6 T. Both magnetic field and radiation beam are orientated perpendicular to the long axis of the alanine holder.

In both ^{60}Co and Elekta Synergy linac radiation beams, the long axis of the alanine holder was positioned in a) the centre of the two magnetic poles, b) perpendicular orientation to the radiation beam, along z-axis and c) perpendicular orientation to the magnetic field, along the x-axis. The magnetic field uniformity in an area of 3 cm² along the centre of the two magnetic poles (in z and y direction) was found to be within ± 0.5 mT. The radiation field was collimated to fill the gap between the two magnetic poles. A calibrated thermistor was placed close to the holder to record the ambient temperature needed to correct the effect of the alanine signal due to temperature.

The effect of the magnetic field on the photon beam symmetry of the linac was checked. A Sun Nuclear IC Profiler was set at 100 cm iso-centre (using the same gantry angle as the experimental setup) and irradiated with a 10 cm x 10 cm beam field. With a magnetic flux density of 0 T, initial beam steering adjustment was performed to give zero tilt using the IC Profiler. The 2T (crossline) and 2R (inline) error values, which indicate if there is any asymmetry of the beam, were set to zero. The magnetic flux density was set to 1.5 T and it was found that symmetry was different by 1.5% compared to profiles without magnetic field (crossline and Inline profiles with magnetic flux densities set at 0 T and 1.5 T are shown on Figure 3). Beam steering was adjusted to eliminate the tilt (2T and 2R error values set back to zero) and the profile symmetry matched the beam without magnetic field. Therefore, it was concluded that zeroing the 2T and 2R error values was a good method to restore the symmetry of the photon beam when the magnetic field was on.




Figure 3: Inline and crossline profiles with 1.5 T and without magnetic field of a 10 cm x 10 cm at 8 MV linac beam.

Here, we must point out, that the magnetic flux density at the iso-centre was less than 0.05 mT, when the electromagnet was set up to give 1.5 T between the poles, and any effect on the IC profiler signal due to the magnetic field was found to be negligible.

3.4 AIR GAPS ON THE FARMER COMPATIBLE ALANINE HOLDER

Alanine is a solid-state detector and is expected not to have a similar strong effect on ERE as air filled ionisation chambers. However, air gaps exist inside the holder where the alanine pellets are located, and two cases of air gaps were identified and investigated:

- 1. Bevel-air-gap: alanine pellets do not have a flat surface, but instead they form a bevelled edge (Figure 4), which introduces air gaps when loaded in the holder. This creates a maximum of 0.052 cm air gap around the bevelled edge between the two pellets. The first (close to the tip) and the last (close to the stem) pellet form the half of this air gap (0.026 cm).
- 2. Cylindrical-air-gap: the internal diameter of the holder has a tolerance of 0.02 cm to allow the pellets to be loaded and unloaded in the holder. This forms an air gap between the inner holder wall and the pellets of 0.01 cm (Figure 4).

MC simulations were performed for both cases to investigate the effect of air gaps on the alanine signal. This effect must be included as a component in the measurement uncertainty, when alanine is used as a dosimeter in the presence of the magnetic field. This is further explained in section 4.8.2. In any case, any effect (including the fluence perturbation by the PEEK holder) will be part of the alanine magnetic field correction factor, explained in section 3.2.




Figure 4: Alanine bevel-air-gap of 0.052 cm, formed due to the pellet's bevelled edge, and cylindrical-air-gap of 0.01 cm, due to a slightly bigger internal diameter of the holder compared to the external diameter of the alanine pellet.

3.5 MONTE CARLO SIMULATIONS

MC simulations of detector response play a major and increasingly important role in accurate radiation dosimetry, including the characterisation of physical properties of beams and of radiation detectors. MC radiation transport codes, such as EGSnrc and PENELOPE have demonstrated a self-consistency test, by performing a Fano cavity test on MV beams, at a level of 0.1%. It has also been shown that the level of agreement with ionisation chamber experimental data, being the most difficult challenge for such a technique, is of the order of 0.3%.

In the presence of the magnetic field, studies (Bouchard *et al.*, 2015, Bouchard and Bielajew, 2015, de Pooter *et al.*, 2015) have demonstrated that it is necessary to define special conditions under which the Fano cavity test can be performed to check the consistency of MC transport algorithms. One of the key transport parameters in EGSnrc that needs to be adjusted to ensure more accurate MC simulations in the presence of magnetic field, is EM ESTEPE. This parameter sets a value for an additional step size restriction based on magnetic field change and direction change for charged particle transport in an electromagnetic field. Research undertaken in the EMPIR project of MR-guided radiotherapy (EMPIR 15HLT08 MRgRT), has shown that the EM ESTEPE parameter needs to be reduced from its default value when the material density is low (i.e. for air the EM ESTEPE = 0.01) in order to achieve the desired level of internal consistency of around 0.1% with EGSnrc (MRgRT, 2019).

In this study, MC simulations were performed to include calculations of absorbed dose to water and to alanine, the investigation on the effect of the air gaps around the alanine pellets and the uncertainty estimation due to the random position of the alanine pellets inside the holder.

The usercode cavity that forms part of the EGSnrc code system (Kawrakow *et al.*, 2011) (development version: GitHub: Aug 2017) was used. An accurate model of the experimental setup consisting of the two magnetic poles, the bespoke water tank, the holder and the alanine pellets was first constructed. The dimensions and material specifications used were taken from the associated manuals, test certificates and in-house dimensional measurements. In the MC simulation, the pellet medium was set to be a 90:10 mixture of alanine and paraffin wax binder. The medium density was set to the measured pellet bulk density (1.23 g/cm³), and the alanine stopping power density effect correction was based on the crystalline density (1.42 g/cm³). Following Zeng *et al.* (2005) the alanine/EPR signal was replaced by dose to the pellet mixture since, over the relevant range of electron energy, variations in the paraffin/alanine stopping power ratio are confined within a range of

$\pm 0.15\%$. The effect of this approximation on the calculation of the dose to alanine ratio, with and without the magnetic field, is expected to be negligible.

The cut-off energy for electrons (AE and ECUT) was set to 0.521 MeV and for photons (AP and PCUT) to 0.01 MeV. For the simulations of the external magnetic field, the default electromagnetic field macros (EMF_MACROS) were enabled, and EM ESTEPE was set to 0.01. All other parameters were set to their default values.

Phase space data of the 6 MV and 8 MV Elekta Synergy x-ray beams were generated at 270 cm from the target, using the BEAMnrc usercode, for the purpose of the detector simulations. For beam validation at 0 T, depth dose and profile measurements were performed at 279 cm from the target. The radiation beam was collimated, at the machine isocentre, by 1.9 cm and 4.4 cm in the crossline and inline direction, respectively. A parallel plate chamber (NACP-02) was placed in a water tank for depth dose measurements, and profiles were measured by irradiating EBT-3 films with a 5 cm build-up. Films were processed and analysed using a method developed by Bouchard *et al.* (2009). A model of a previously validated ^{60}Co beam (Costa *et al.*, 2018), which configures the same beam characteristics as in this study, was used for the MC simulations.

The following three sections (3.5.1, 3.5.2, 3.5.3) will describe the MC simulations that were performed to support investigations in this study.

3.5.1 Experimental setup model validation

For the model validation of the experimental setup, including the electromagnet, alanine and holder, measurements were performed using the setup explained in section 3.3. In this case, measurement and MC simulations were performed with the holder partially loaded with alanine pellets, so that an air gap of nominal 0.44 cm occurs between the stem and the first pellet (Figure 5). In this air gap, the electrons follow on average a curved trajectory in accordance with the Lorentz force by increasing the dose on the first pellet, which are distributed along the long axis of the holder towards the tip. The magnetic flux density was 1.5 T and alanine pellets were irradiated with 6 MV and 8 MV delivering 20 Gy in each case. In the simulations, the total dose per incident particle was scored in each pellet. Unfortunately, it was not possible to perform similar measurements to validate the experimental setup at 1.5 T in the ^{60}Co beam during the study. Nevertheless, the characteristics of the ^{60}Co beam model, used in this study, have been benchmarked by Costa *et al.* (2018).




Figure 5: Alanine holder filled with pellets for the purpose of experimental setup validation. The direction of the magnetic field (pointing into the plane of the paper) and radiation beam (lying in the plane of the paper) with respect to the holder are shown together with the affected electron trajectories.

3.5.2 Monte Carlo model for the air gap effect

MC simulations were performed to determine the effect of the air gaps around the alanine pellets. A model of the holder loaded with alanine pellets was used and three sets of simulations were performed: a) full model including both air gaps (bevel and cylindrical), b) full model with the air gaps created from the bevelled edge of the alanine pellets (bevel-air-gaps only) filled with alanine medium and c) full model with no air gaps (both air gaps filled with alanine medium). Figure 6 depicts the three models. Simulations were performed for all beam qualities at all examined magnetic flux densities. For each combination, the total dose per incident particle was scored to the first five pellets from the tip of the holder. Any effect of the magnetic field on the dose distribution in alanine medium was removed (for 1.5 T the dose was found to be negatively deviated by 0.03%, 0.3% and 0.4% from the dose at 0 T for ^{60}Co , 6 MV and 8 MV, respectively). Thus, only the air gap effect is considered for the comparison.




Figure 6: MC models for the investigation of the effect of the air gaps around the alanine pellets.

3.5.3 Uncertainty due to the air gaps

When alanine pellets are loaded inside the holder, they will not be located symmetrically but rather randomly due to the air gap of 0.02 cm (see section 3.4). Depending on the position of the air gaps inside the holder with respect to the magnetic field, the ERE will either increase or decrease the dose in each alanine pellet. Due to the unpredictable location of the pellets, it is difficult to calculate what the effect would be and correct for it. However, known air gaps of 0.02 cm, which is the maximum between the holder and the alanine pellets, can be modelled by MC simulations to estimate the effect and include it in the uncertainties (see section 4.8.2). For that purpose, a model of the holder loaded with nine alanine pellets was built and MC simulations were performed by shifting the pellets in the x- and z-axis as shown in Figure 7 (dimensions not to scale), allowing the maximum achievable air-gap (0.02 cm). The total dose per incident particle was scored to the first five pellets from the tip of the holder for each of the three beam energies, for all magnetic flux densities and all five combinations.




Figure 7: MC models to evaluate uncertainties due to air gaps. (a) coordinates indicating the direction of magnetic field and radiation beam with respect to the alanine holder, (b) model of the centred alanine pellets in the holder and (c, d, e, f) offset alanine pellets along the short axis (x- and z-axis) creating an air gap of 0.02 cm. Dimensions not to scale.

4 RESULTS

4.1 VALIDATION OF THE BEAM MODELS

The 6 MV and 8 MV beam models were validated by comparing the simulated with the experimental dose profiles and depth doses as shown in Figure 8. For each plot, the data were normalised to the integral of the curve to minimise the effect of point-to-point variations in film sensitivity. The MC profiles agree with the film measurements within the uncertainty of the single channel analysis made of the film data, which in this study was estimated to be 2%.




Figure 8. Validation of the MC models by comparing experimental and simulated profiles and depth doses, at 1.5 T, for 6 MV and 8 MV beam energies.

4.2 ALANINE MODEL AND EXPERIMENTAL SET UP VALIDATION IN A MAGNETIC FIELD

The validation of the experimental setup model for 6 MV and 8 MV with experiments was performed at a magnetic flux density of 1.5 T. It was found that the thickness of the air gap between the PEEK stem and the first alanine pellet, depends on the force that is applied to adjust the screw cap shown in Figure 1, which is used to secure the pellets inside the holder. Further measurements of five actual holders and pellets showed typical variations of the air gap in the order of 0.35 ± 0.02 cm, and so additional simulations with air gaps of 0.30 cm and 0.40 cm were made. Figure 9 shows the MC results (for three different air gaps) and the measurement results of the dose deposited at each alanine pellet, normalised to the average dose of all pellets, for 6 MV and 8 MV. The experimental data points for both energies agree with the MC calculated points of the two additional simulations, of 0.30 cm and 0.40 cm air gaps, within the measurement uncertainties. The error bars on the experimental data represent the combined standard uncertainty of an alanine pellet (0.6%) and the uncertainty due to the air gaps as explained in section 4.8.2. The uncertainty on the MC data is in the order of 0.15%. For clarity, error bars are not included on the MC data.




Figure 9. MC and measurement results, including air gaps of 0.30 cm, 0.40 cm and 0.44 cm (nominal air gap), of the dose deposited at each alanine pellet, normalised to the average dose of all pellets, for 6 MV and 8 MV, plotted against the distance measured from the front of the first alanine pellet. Note that the air gaps, on the x-axis of the two plots, are at the front face (0 cm).

4.3 EFFECT OF AIR GAPS ON ALANINE IN A MAGNETIC FIELD

Figure 10 shows the effect of the air gaps on the calculated dose-to-alanine for the three beam energies and all tested magnetic flux densities, based on the three sets of simulations described in section 3.5.2. Calculated data show how the dose to alanine changes when the air gaps created from the bevelled edge are simulated with alanine replacing the air, Figure 10 (a), and when both bevel and cylindrical air gaps are replaced with alanine medium, i.e. full model with no air gaps, Figure 10 (b). In both cases, data were normalised to the MC results of the full model (including both air gaps). In Figure 10 (b), the maximum deviation from the full model was found to be 0.45% and 0.55% for 6 MV and 8 MV, respectively, at 1.5 T. In Figure 10 (a), the maximum deviation was found to be at 1 T and 1.5 T and was less than 0.4% for both 6 MV and 8 MV. For the ⁶⁰Co beam all data from both figures were found to be less than 0.4%.




Figure 10: MC results of the effect of the air gaps on the calculated dose on alanine for ⁶⁰Co, 6 MV and 8 MV at all tested magnetic flux densities, when the bevelled-air-gaps (a) and the annular-air-gap (b) are filled with alanine medium. Data are normalised to the MC results of full model. Error bars represent the combined MC uncertainties.

4.4 EFFECT OF THE MAGNETIC FIELD ON ABSORBED DOSE TO WATER

The effect of the magnetic field on absorbed dose to water at the measurement point, $D_{w,Q}$, is needed to determine the alanine magnetic field quality correction factor and its relative intrinsic sensitivity, as explained in section 3.2. For that purpose, depth doses in water for each beam energy and at magnetic flux densities up to 3 T were calculated by using MC simulations. Figure 11 shows the MC calculated $D_{w,Q_B}/D_{w,Q}$ ratios as a function of magnetic flux density for the three beam energies. After d_{max} , for all beam energies, depth doses in the presence of a magnetic field, although systematically lower compared to 0 T, were found to be similar. Therefore, the $D_{w,Q_B}/D_{w,Q}$ ratios at each magnetic flux density were determined from the average ratios after d_{max} , which results in a level of standard uncertainty of 0.04%.




Figure 11: MC calculated $D_{w,Q_B}/D_{w,Q}$ ratios as a function of magnetic flux density for ^{60}Co , 6 MV and 8 MV.

4.5 ALANINE RESPONSE IN A MAGNETIC FIELD

Figure 12 summarises the alanine/EPR signal for measurements in a magnetic field, M_{al,Q_B} , normalised to the signal at zero magnetic field, $M_{al,Q}$, as a function of magnetic flux density for the three beam energies. For each plot and each magnetic flux density, data represent measurements over three days and their distribution is displayed in a form of box plot. Note that crosses (x) on each box plot indicate the mean value of the normalised alanine signal for each magnetic flux density over three days.

For zero magnetic field and for each energy, the pellet-to-pellet variation of the signal has a standard deviation of up to 0.52%. This results in a standard uncertainty of the mean of 0.13% or better, considering the number of the pellets used in this study. However, as the magnetic flux density is increased, for each energy, the standard deviation also increases, up to 0.85% (standard uncertainty of the mean of 0.22%).




Figure 12: Box plots presenting the ratio between the alanine/EPR signal for alanine measurements in a magnetic field, M_{Al,Q_B} , and without magnetic field, $M_{Al,Q}$, for ^{60}Co , 6 MV and 8 MV. Data at each energy and each magnetic flux density represent measurements over three days. Crosses (x) show the mean value of the normalised alanine signal for each magnetic flux density over three days.

4.6 QUALITY CORRECTION FACTOR FOR THE PRESENCE OF A MAGNETIC FIELD ON ALANINE

The alanine quality correction factor as a function of a magnetic flux density is shown in Figure 13 and Table 1, for the three beam energies. The error bars denote the overall combined standard uncertainty (ranging from 0.2% to 0.6%), as quoted in section 4.8, and include the effects of measurement repeatability, air gaps, linac drift and MC uncertainties. Figure 13 shows how the correction factor for each energy decreases with increasing magnetic flux density. Except for the data point at 0.5 T (1.0014) for 8 MV, all the remaining data points lie below unity, ranging from 0.9933 to 0.9998 with an average of 0.9967 ± 0.0027 . Note that this correction includes both the effect of the magnetic field on absorbed dose to water and its effect on alanine intrinsic sensitivity. The latter effect is isolated in the following.




Figure 13: Alanine quality correction factor for the presence of magnetic field, $k_{Q_B,Q}$, for all examined magnetic flux densities for ^{60}Co , 6 MV and 8 MV. The error bars denote the standard combined uncertainties estimated in section 4.8.

Table 1: Alanine quality correction factor for the presence of a magnetic field, $k_{Q_B,Q}$, for all examined magnetic flux densities for the three beam energies. Standard uncertainties estimated in section 4.8.

Magnetic flux density (T)	^{60}Co	6 MV	8 MV
0.35	-	0.9992 ± 0.0026	0.9976 ± 0.0027
0.50	0.9974 ± 0.0024	0.9984 ± 0.0027	1.0014 ± 0.0026
1.00	0.9940 ± 0.0027	0.9954 ± 0.0040	0.9998 ± 0.0034
1.50	0.9939 ± 0.0034	0.9964 ± 0.0060	0.9938 ± 0.0055
2.00	0.9933 ± 0.0039	-	-

4.7 THE EFFECT OF THE MAGNETIC FIELD ON ALANINE RESPONSE

The magnetic field effect on alanine intrinsic sensitivity may be represented by the quantity $F_{Q_B,Q}$, relative intrinsic sensitivity, which is defined as the ratio of the alanine intrinsic sensitivity with and without magnetic field (equations (5) and (6) in section 3.2). This was obtained by combining measurements with MC simulations and is shown in Figure 15 and Table 2, for the three beam energies. The alanine over-response tends to increase with magnetic flux density, up to 1.0099 at 8 MV in 1.5 T, but without any obvious trend as a function of energy.




Figure 14: Alanine relative intrinsic sensitivity, $F_{Q_B, Q}$, for all examined magnetic flux densities for ^{60}Co , 6 MV and 8 MV. The error bars denote the standard combined uncertainties estimated in section 4.8.

Table 2: Alanine relative intrinsic sensitivity, $F_{Q_B, Q}$, for all examined magnetic flux densities for the three beam energies. Standard uncertainties estimated on section 4.8.

Magnetic flux density (T)	^{60}Co	6 MV	8 MV
0.35	-	1.0011 ± 0.0030	1.0038 ± 0.0030
0.50	1.0035 ± 0.0027	1.0030 ± 0.0031	1.0006 ± 0.0030
1.00	1.0079 ± 0.0030	1.0082 ± 0.0042	1.0057 ± 0.0036
1.50	1.0051 ± 0.0037	1.0067 ± 0.0061	1.0099 ± 0.0056
2.00	1.0050 ± 0.0042	-	-

4.8 UNCERTAINTIES

The analysis of uncertainty here follows the Joint Committee for Guides in Metrology (JCGM) Guide to the Expression of Uncertainty in Measurement (JCGM, 2008). Uncertainties evaluated by statistical analysis are grouped as type A and the rest are grouped as type B. These are added in quadrature to give a combined standard uncertainty with coverage factor $k = 1$.

The quoted relative standard uncertainties are shown in Table 3. **Error! Reference source not found..** This includes the overall combined relative standard uncertainties in $k_{Q_B, Q}$ at 0.35 T and 1.5 T for 6 MV and 8 MV. For ^{60}Co , uncertainties are presented at 0.5 T and 1.5 T. For the rest of the magnetic flux densities at all three energy beams, the relative standard uncertainties are between the stated uncertainties in Table 3. For ^{60}Co and 2 T, the uncertainty was found to be 0.56%.

Table 3: Uncertainty budget for the determination of the quality correction factor for the presence of magnetic field on alanine, $k_{Q_B,Q}$.

Uncertainty component	Type	Relative standard uncertainty (%)					
		^{60}Co		6 MV		8 MV	
		0.5 T	1.5 T	0.35 T	1.5 T	0.35 T	1.5 T
Measurement repeatability	A	0.06	0.22	0.06	0.20	0.08	0.11
Air gaps	B	0.11	0.29	0.13	0.52	0.19	0.47
Linac output correction	A	-	-	0.07	0.07	0.10	0.10
Monte Carlo	A+B	0.15	0.15	0.15	0.15	0.15	0.15
Overall combined relative standard uncertainty on $k_{Q_B,Q}$		0.20	0.39	0.22	0.58	0.27	0.52

4.8.1 Measurement uncertainty and repeatability

The uncertainty on the repeatability of the measurements is evaluated by considering the behaviour of the averaged signal over the alanine pellets at each magnetic flux density for each energy, over all experiments. The standard deviation of the mean was found to be ranging from 0.05% to 0.22%.

4.8.2 Uncertainty due to air gaps

Uncertainties due to the unknown spatial distribution of the alanine pellets in the PEEK holder were estimated using MC simulations as explained in section 3.5.3. Figure 15 (a) shows the MC calculated dose to alanine for each energy, when pellets are shifted in the x- and z-axis inside the alanine holder (see Figure 6). For clarity, Figure 15 (a), only includes data for 1.5 T, which are normalised with respect to when the pellets are centrally-located in the holder. This figure shows how the ERE is varying the dose to alanine at the different positions. The type A uncertainties were based on the root mean square (RMS) error of the four different distributions of the alanine pellets inside the holder. The uncertainties, for each energy, were found to increase with magnetic flux density up to 0.52% and 0.47% for 6 MV and 8 MV at 1.5 T, respectively. For ^{60}Co it was 0.52% at 2 T. For the other magnetic flux densities and for all beam energies the determined uncertainties were below 0.30%. The rise in uncertainties as a function of magnetic flux density is reflected by the length of the error bars in Figure 15 (b), (c) and (d). These figures show the average normalised dose over the four shifted alanine positions, as per Figure 15 (a), for each energy beam at all examined magnetic flux densities. It can be observed that the error bars, at each energy beam, increases with magnetic flux density. The uncertainty due to the air gaps is the dominant component in the uncertainty budget and is applicable to the absorbed dose to water measured with the alanine detector, in the presence of a magnetic field using the holder designed for this study.




Figure 15: MC simulations to investigate the uncertainty estimation due to the random position of the alanine pellets inside the PEEK holder. The MC calculated dose to alanine for all examined magnetic flux densities, when pellets are shifted in four different positions, with respect to the x- and z-axis inside the alanine holder, are shown for ^{60}Co , 6 MV and 8 MV at 1.5 T (a). All data are normalised with respect to when the pellets are centrally-located in the holder. The average normalised dose over the four shifted positions, at the examined magnetic flux densities, are shown for ^{60}Co (b), 6 MV (c) and 8 MV (d).

4.8.3 Uncertainty due to linac output

The short term (over one day) behaviour of the output from the Elekta Synergy linac beams was investigated during the alanine irradiation. The output was measured several times in-between the alanine irradiations using an ionisation chamber with the same experimental set up and always at 0 T. For each day of irradiation, the output (dose rate) was found to diverge by an average of -0.07% per hour. This deviation was considered, and used to correct the alanine signal, based on a linear fit between the ionisation chamber signal and the time. The uncertainty was estimated from the gradient of the residuals of the fit (RMS error) and estimated to be less than 0.1%, on average.

4.8.4 Monte Carlo uncertainties

In the MC simulations, the Type A uncertainties in the determination of the absorbed dose to water, and of the absorbed dose to alanine, with and without magnetic field, were typically 0.1%. For self-consistency and transport parameters we estimated a Type B uncertainty of 0.1% (Kawrakow, 2000, Malkov and Rogers, 2016). The combined standard uncertainty on the MC simulations resulted in 0.15%.

5 DISCUSSION

The performance of the alanine dosimeter in the presence of a magnetic field has been investigated at three different photon beam energies over a range of magnetic flux densities. Alanine pellets were placed in a waterproof PEEK holder, developed in-house and shaped to match a Farmer-type chamber, in an electromagnet and irradiated by either a ^{60}Co source or conventional 6 MV and 8 MV Elekta Synergy linac beams. The long axis of the alanine holder was positioned in the centre of the magnetic poles of the electromagnet and in perpendicular orientation relative to both radiation beam and the magnetic field.

5.1 MC SIMULATIONS AND AIR GAP EFFECTS

MC simulations were performed to: a) investigate the effect of the air gaps on dose to alanine (found around the pellets when loaded in the PEEK holder), b) investigate the uncertainty due to the random position of the pellets inside the holder and c) calculate the absorbed dose to water and absorbed dose to alanine at the three beam qualities, with and without magnetic field, for the determination of the quality correction factor for the presence of magnetic field on alanine.

The linac 6 MV and 8 MV photon beam MC models were validated by comparing simulated with experimental beam profiles and depth doses. For both energies a very good agreement was found between the simulations and the measurements. In validating the MC simulations of the experimental set up, for both energies at 1.5 T, the holder was loaded with alanine pellets, creating an air gap between the holder's stem and the first pellet. It became apparent that the thickness of the air gap varies with respect to the force that is applied on the screw cap (Figure 1) to secure the pellets in the holder. Thus, MC simulations were performed for different air gaps and it was demonstrated that at least two of the MC models reproduced well the deviated behaviour of the dose along the pellets towards the tip of the holder.

The results of the MC simulations show that the air gaps affect the alanine response, due to the ERE caused by the magnetic field, by as much as 0.45% and 0.55% for 6 MV and 8 MV, respectively, and less than 0.40% for ^{60}Co over all the magnetic flux densities. The air gaps around other dosimeters, such as ionisation chambers, will increase the variation on the dose measurement, which can be as much as 3.8% for a Farmer-type chamber (Agnew *et al.*, 2017). This can be eliminated by immersing the dosimeters in water (if waterproof). Alanine, however, is not waterproof and needs to be placed in a watertight holder for dose measurements in water. Due to the structure of the alanine pellets (they form bevelled edges) and the possible asymmetry in the positions of them inside the holder, it is difficult to avoid the effect caused by the existing air gaps. However, we can include this effect as a component in the measurement uncertainty, which increases with magnetic flux density and reaches 0.52% for 6 MV and 0.47% for 8 MV at 1.5 T. For ^{60}Co the highest uncertainty was 0.52% at 2 T (0.29% at 1.5 T).

The maximum pellet-to-pellet variation of the measured alanine EPR/signal at 0 T, based on the number of the pellets considered in this study for the three energies, was found to be 0.52%, which results in a standard uncertainty of the mean of 0.13% or less. This is in line with previous studies (McEwen M. *et al.*, 2015, Sharpe and Sephton, 2006, Anton, 2006), which investigated alanine dosimetry at therapy level energy beams. Nevertheless, for each energy, the variation was found to increase with the magnetic flux density up to 0.85%, reaching a standard uncertainty of the mean of 0.22% at 1.5 T. The increase in variation is mainly attributable to the air gaps, which vary in size, based on the random locations of the pellets inside the holder. In this case, the ERE will diverge the alanine EPR/signal, as the path length of the secondary electrons will follow a curved trajectory due to the Lorentz force. These affected electrons will either deposit their energy to alanine or will curve back 'exiting' the holder.

5.2 QUALITY CORRECTION FACTOR FOR THE PRESENCE OF A MAGNETIC FIELD ON ALANINE

The alanine quality correction factor for the presence of a magnetic field was determined by combining measurements of the alanine/EPR signal with calculations of the absorbed dose, to water and to alanine, in the presence of a magnetic field, by MC simulations. Figure 13 shows that the correction factor tends to decrease with increasing magnetic flux density. For each energy the correction, averaged over all non-zero magnetic flux densities, is 0.9946 ± 0.0019 for ^{60}Co , 0.9973 ± 0.0018 for 6 MV and 0.9982 ± 0.0033 for 8 MV. Although the change in the average correction factor with energy is monotonic, the uncertainties are too large to allow a more precise conclusion regarding energy dependence. The overall correction factor, $k_{QB,Q}$, takes account of all the effects of the magnetic field, on dose to water, dose to alanine, alanine sensitivity, and the fluence perturbation by the PEEK holder and by the air gaps. These effects are built in the $k_{QB,Q}$, which can be used to correct the dose to water measured with alanine in the presence of a magnetic field, when alanine is calibrated in the absence of a magnetic field. If a different holder was used (holder material, air gap details, etc.), it would be necessary to recalculate the $k_{QB,Q}$ correction factor. This correction factor is holder-dependent.

5.3 THE EFFECT OF THE MAGNETIC FIELD ON ALANINE RESPONSE

The effect on the alanine intrinsic sensitivity due to the magnetic field, $F_{QB,Q}$, can be calculated by using the experimental data and the MC simulations as defined in equation (6). The intrinsic sensitivity describes the yield of stable free radicals per unit dose to alanine. It can be seen from Figure 14, that the alanine intrinsic sensitivity increases with magnetic flux density for all examined beam energies, without any obvious energy dependence. All $F_{QB,Q}$ values lie above unity with the highest being 1.0099 ± 0.0056 at 8 MV and at 1.5 T. A common linear fit $F_{QB,Q} \sim (1 + aB)$ was made to the data, for all energies, and this indicates that the effect of a magnetic field on alanine is to increase its relative intrinsic sensitivity by $a = 0.0047 \pm 0.0005$ per T. This value, a , may be used to estimate the required correction factor, $k_{QB,Q}$, for different flux densities, B , by combining the fit value of $F_{QB,Q}$ with MC simulations of absorbed dose, to water and to alanine.

6 CONCLUSION

This study successfully characterised the performance of alanine irradiated in the presence of a magnetic field. It was found that the effect of the magnetic field on an alanine/EPR signal is energy independent and may be increased by 0.2% at 0.35 T and 0.7% at 1.5 T. Neglecting this effect, may introduce systematic errors in the measured absorbed dose to water in commercially available MRI linacs. Therefore, if alanine is calibrated in the absence of any magnetic field, a correction factor, $k_{QB,Q}$, needs to be applied for the determination of the true absorbed dose to water in the presence of a magnetic field. This correction factor, which depends on the geometry and material of the alanine holder, varies with magnetic flux density. Not applying this correction factor would results in an overestimation of the dose to water. With the inclusion of $k_{QB,Q}$, alanine as an MR safe, robust and stable dosimeter with comparable uncertainties to a Farmer-type chamber, is a suitable dosimeter to act as a reference class detector for MRI-guided radiotherapy.

7 ACKNOWLEDGMENTS

The work was supported by the UK government's Department for Business, Innovation and Skills and the European Metrology Programme for Innovation and Research (EMPIR) project 15HLT08 MRgRT. The EMPIR initiative is co-funded by the European Union's Horizon 2020 research and innovation programme and the EMPIR participating states. Research at The Institute of Cancer Research is

supported by Cancer Research UK under Programme C33589/A19727 and NHS funding to the NIHR Biomedical Research Centre at The Royal Marsden Hospital and The Institute of Cancer Research.

8 REFERENCES

- Agnew, J., O'Grady, F., Young, R., Duane, S. & Budgell, G. J. 2017. Quantification of static magnetic field effects on radiotherapy ionization chambers. *Physics in Medicine and Biology*, 62, 1731-1743.
- Andreo, P., Burns, D. T., Hohlfield, K., Huq, M. S., Kanai, T., Laitano, F., Smyth, V. & Vynckier, S. 2000. Absorbed dose determination in external beam radiotherapy: an international Code of Practice for dosimetry based on standards of absorbed dose to water. *IAEA Technical Reports Series no 398 (Vienna: International Atomic Energy Agency)*.
- Anton, M. 2006. Uncertainties in alanine/ESR dosimetry at the Physikalisch-Technische Bundesanstalt. *Physics in Medicine and Biology*, 51, 5419-5440.
- Anton, M., Kapsch, R. P., Krauss, A., von Voigts-Rhetz, P., Zink, K. & McEwen, M. 2013. Difference in the relative response of the alanine dosimeter to megavoltage x-ray and electron beams. *Physics in Medicine and Biology*, 58, 3259-3282.
- Bergstrand, E. S., Bjerke, H. & Hole, E. O. 2005. An experimental investigation of the electron energy dependence of the EPR alanine dosimetry system. *Radiation Measurements*, 39, 21-28.
- Bergstrand, E. S., Shortt, K. R., Ross, C. K. & Hole, E. O. 2003. An investigation of the photon energy dependence of the EPR alanine dosimetry system. *Physics in Medicine and Biology*, 48, 1753-1771.
- Bouchard, H. & Bielajew, A. 2015. Lorentz force correction to the Boltzmann radiation transport equation and its implications for Monte Carlo algorithms. *Physics in Medicine and Biology*, 60, 4963-4971.
- Bouchard, H., de Pooter, J., Bielajew, A. & Duane, S. 2015. Reference dosimetry in the presence of magnetic fields: conditions to validate Monte Carlo simulations. *Physics in Medicine and Biology*, 60, 6639-6654.
- Bouchard, H., Lacroix, F., Beaudoin, G., Carrier, J. F. & Kawrakow, I. 2009. On the characterization and uncertainty analysis of radiochromic film dosimetry. *Medical Physics*, 36, 1931-1946.
- Bouchard, H. & Seuntjens, J. 2013. Applications of Monte Carlo to radiation dosimetry. In: SECO, J. & VERHAEGEN, F. (eds.) *Monte Carlo Techniques in Radiation Therapy*. 1 ed. Boca Raton: CRC Press.
- Budgell, G., Brown, K., Cashmore, J., Duane, S., Frame, J., Hardy, M., Paynter, D. & Thomas, R. 2016. IPEM topical report 1: guidance on implementing flattening filter free (FFF) radiotherapy. *Physics in Medicine and Biology*, 61, 35.

- Burns, D. T., Allisy-Roberts, P. J., Desrosiers, M. F., Sharpe, P. H. G., Pimpinella, M., Lourenco, V., Zhang, Y. L., Miller, A., Generalova, V. & Sochor, V. 2011. Supplementary comparison CCRI(I)-S2 of standards for absorbed dose to water in Co-60 gamma radiation at radiation processing dose levels. *Metrologia*, 48, 18.
- Costa, F., Doran, S. J., Hanson, I. M., Nill, S., Billas, I., Shipley, D., Duane, S., Adamovics, J. & Oelfke, U. 2018. Investigating the effect of a magnetic field on dose distributions at phantom-air interfaces using PRESAGE (R) 3D dosimeter and Monte Carlo simulations. *Physics in Medicine and Biology*, 63, 10.
- de Pooter, J. A., de Prez, L. A. & Bouchard, H. 2015. Application of an adapted Fano cavity test for Monte Carlo simulations in the presence of B-fields. *Physics in Medicine and Biology*, 60, 9313-9327.
- de Prez, L., de Pooter, J., Jansen, B., Woodings, S., Wolthaus, J., van Asselen, B., van Soest, T., Kok, J. & Raaymakers, B. 2019. Commissioning of a water calorimeter as a primary standard for absorbed dose to water in magnetic fields. *Physics in Medicine and Biology*, 64, 12.
- Dimitriadis, A., Palmer, A., Thomas, R., Nisbet, A. & Clark, C. 2017. Adaptation and validation of a commercial head phantom for cranial radiosurgery dosimetry end-to-end audit.
- Hackett, S. L., van Asselen, B., Wolthaus, J. W. H., Kok, J. G. M., Woodings, S. J., Lagendijk, J. J. W. & Raaymakers, B. W. 2016. Consequences of air around an ionization chamber: Are existing solid phantoms suitable for reference dosimetry on an MR-linac? *Medical Physics*, 43, 3961-3968.
- JCGM 2008. Evaluation of measurement data – Guide to the expression of uncertainty in measurement. JCGM 100:2008 (GUM 1995 with minor corrections).
- Kawrakow, I. 2000. Accurate condensed history Monte Carlo simulation of electron transport. II. Application to ion chamber response simulations. *Medical physics*, 27, 499-513.
- Kawrakow I, Mainegra-Hing E, Rogers D W O, Tessier F & B, W. B. R. 2011. The EGSnrc code system: Monte Carlo simulation of electron and photon transport National Research Council of Canada (NRCC) Report PIRS-701.
- Malinen, E. 2014. EPR Dosimetry in Clinical Applications. In: LUND, A. & SHIOTANI, M. (eds.) *Applications of EPR in Radiation Research*. 1 ed.: Springer International Publishing.
- Malkov, V. & Rogers, D. 2016. Charged particle transport in magnetic fields in EGSnrc. *Medical physics*, 43, 4447-4458.

- Malkov, V. N. & Rogers, D. W. O. 2017. Sensitive volume effects on Monte Carlo calculated ion chamber response in magnetic fields. *Medical Physics*, 44, 4854-4858.
- Massillon, G., Cueva-Procel, D., Diaz-Aguirre, P., Rodriguez-Ponce, M. & Herrera-Martinez, F. 2013. Dosimetry for Small Fields in Stereotactic Radiosurgery Using Gafchromic MD-V2-55 Film, TLD-100 and Alanine Dosimeters. *Plos One*, 8, 8.
- McEwen, M. & Ross, C. 2009. Fricke and alanine dosimeters. In: ROGERS, D. W. O. & CYGLER, J. E. (eds.) *Clinical Dosimetry Measurements in Radiotherapy*. Madison, WI: Medical Physics.
- McEwen M., Sharpe P. & Vörös S. 2015. Evaluation of alanine as a reference dosimeter for therapy level dose comparisons in megavoltage electron beams. *Metrologia*, 52, 272-279.
- Meijsing, I., Raaymakers, B. W., Raaijmakers, A. J. E., Kok, J. G. M., Hogeweg, L., Liu, B. & Lagendijk, J. J. W. 2009. Dosimetry for the MRI accelerator: the impact of a magnetic field on the response of a Farmer NE2571 ionization chamber. *Physics in Medicine and Biology*, 54, 2993-3002.
- MRgRT 2019. Metrology for MR guided Radiotherapy. Report retrieved from: https://www.euramet.org/research-innovation/search-research-projects/details/project/metrology-for-mr-guided-radiotherapy/?tx_eurametctcp_project%5Baction%5D=show&tx_eurametctcp_project%5Bcontroller%5D=Project&L=0&cHash=6fe39b78a51062b1168d0d0a12feb6d8.
- O'Brien, D. J., Roberts, D. A., Ibbott, G. S. & Sawakuchi, G. O. 2016. Reference dosimetry in magnetic fields: formalism and ionization chamber correction factors. *Medical Physics*, 43, 4915-4927.
- Pojtinger, S., Dohm, O. S., Kapsch, R. P. & Thorwarth, D. 2018. Ionization chamber correction factors for MR-linacs. *Physics in Medicine and Biology*, 63, 8.
- Pojtinger, S., Kapsch, R. P., Dohm, O. S. & Thorwarth, D. 2019. A finite element method for the determination of the relative response of ionization chambers in MR-linacs: simulation and experimental validation up to 1.5 T. *Physics in Medicine and Biology*, 64, 7.
- Raaijmakers, A. J. E., Raaymakers, B. W. & Lagendijk, J. J. W. 2005. Integrating a MRI scanner with a 6 MV radiotherapy accelerator: dose increase at tissue-air interfaces in a lateral magnetic field due to returning electrons. *Physics in Medicine and Biology*, 50, 1363-1376.
- Raaijmakers, A. J. E., Raaymakers, B. W., van der Meer, S. & Lagendijk, J. J. W. 2007. Integrating a MRI scanner with a 6 MV radiotherapy accelerator:

- impact of the surface orientation on the entrance and exit dose due to the transverse magnetic field. *Physics in Medicine and Biology*, 52, 929-939.
- Raaymakers, B. W., Raaijmakers, A. J. E., Kotte, A., Jette, D. & Lagendijk, J. J. W. 2004. Integrating a MRI scanner with a 6 MV radiotherapy accelerator: dose deposition in a transverse magnetic field. *Physics in Medicine and Biology*, 49, 4109-4118.
- Ramirez, J. L. V., Chen, F., Nicolucci, P. & Baffa, O. 2011. Dosimetry of small radiation field in inhomogeneous medium using alanine/EPR minidosimeters and PENELOPE Monte Carlo simulation. *Radiation Measurements*, 46, 941-944.
- Reynolds, M., Fallone, B. G. & Rathee, S. 2013. Dose response of selected ion chambers in applied homogeneous transverse and longitudinal magnetic fields. *Medical Physics*, 40, 7.
- Sharpe, P. H. G. 2003. Report on Radiation Dosimetry at NPL. Report No. CCRI(I)/03-14.
- Sharpe, P. H. G., Rajendran, K. & Sephton, J. P. 1996. Progress towards an alanine/ESR therapy level reference dosimetry service at NPL. *Applied Radiation and Isotopes*, 47, 1171-1175.
- Sharpe, P. H. G. & Sephton, J. 2000. An automated system for the measurement of alanine/EPR dosimeters. *Applied Radiation and Isotopes*, 52, 1185-1188.
- Sharpe, P. H. G. & Sephton, J. 2006. Therapy level alanine dosimetry at the NPL. *Proceedings of 216th PTB Seminar "Alanine Dosimetry for Clinical Application", PTB, Braunschweig, PTB-Dos-51.*
- Smit, K., van Asselen, B., Kok, J. G. M., Aalbers, A. H. L., Lagendijk, J. J. W. & Raaymakers, B. W. 2013. Towards reference dosimetry for the MR-linac: magnetic field correction of the ionization chamber reading. *Physics in Medicine and Biology*, 58, 5945-5957.
- Spindeldreier, C. K., Schrenk, O., Bakenecker, A., Kawrakow, I., Burigo, L., Karger, C. P., Greilich, S. & Pfaffenberger, A. 2017. Radiation dosimetry in magnetic fields with Farmer-type ionization chambers: determination of magnetic field correction factors for different magnetic field strengths and field orientations. *Physics in Medicine and Biology*, 62, 6708-6728.
- Thomas, R. A. S., Duane, S., McEwen, M. R. & Rosser, K. E. 2003. *Role of the National Physical Laboratory in monitoring and improving dosimetry in radiotherapy in the United Kingdom*, International Atomic Energy Agency (IAEA), IAEA.
- van Asselen, B., Woodings, S. J., Hackett, S. L., van Soest, T. L., Kok, J. G. M., Raaymakers, B. W. & Wolthaus, J. W. H. 2018. A formalism for reference

dosimetry in photon beams in the presence of a magnetic field. *Physics in Medicine and Biology*, 63, 9.

Zeng, G. G., McEwen, M. R., Rogers, D. W. O. & Klassen, N. V. 2004. An experimental and Monte Carlo investigation of the energy dependence of alanine/EPR dosimetry: I. Clinical x-ray beams. *Physics in Medicine and Biology*, 49, 257-270.

Zeng, G. G., McEwen, M. R., Rogers, D. W. O. & Klassen, N. V. 2005. An experimental and Monte Carlo investigation of the energy dependence of alanine/EPR dosimetry: II. Clinical electron beams. *Physics in Medicine and Biology*, 50, 1119-1129.