

Accepted Manuscript

Combined ultrasound and cone beam CT improves target segmentation for image guided radiation therapy in uterine cervix cancer

Sarah A. Mason, PhD, Ingrid M. White, FRCR, Tuathan O'Shea, PhD, Helen A. McNair, PhD, Sophie Alexander, MSc, Eleftheria Kalaitzaki, MSc, Jeffrey C. Bamber, PhD, Emma J. Harris, PhD, Susan Lalondrelle, FRCR

PII: S0360-3016(19)30316-5

DOI: <https://doi.org/10.1016/j.ijrobp.2019.03.003>

Reference: ROB 25584

To appear in: *International Journal of Radiation Oncology • Biology • Physics*


Received Date: 13 November 2018

Revised Date: 7 February 2019

Accepted Date: 3 March 2019

Please cite this article as: Mason SA, White IM, O'Shea T, McNair HA, Alexander S, Kalaitzaki E, Bamber JC, Harris EJ, Lalondrelle S, Combined ultrasound and cone beam CT improves target segmentation for image guided radiation therapy in uterine cervix cancer, *International Journal of Radiation Oncology • Biology • Physics* (2019), doi: <https://doi.org/10.1016/j.ijrobp.2019.03.003>.

This is a PDF file of an unedited manuscript that has been accepted for publication. As a service to our customers we are providing this early version of the manuscript. The manuscript will undergo copyediting, typesetting, and review of the resulting proof before it is published in its final form. Please note that during the production process errors may be discovered which could affect the content, and all legal disclaimers that apply to the journal pertain.



Full title: Combined ultrasound and cone beam CT improves target segmentation for image guided radiation therapy in uterine cervix cancer

Short running title: Target localization with CBCT-US fusion

Authors:

Sarah A. Mason, PhD, Institute of Cancer Research, Radiotherapy and Imaging, London, UK

Ingrid M. White, FRCR, Royal Marsden NHS Foundation Trust, Radiotherapy Department, London, UK

Tuathan O'Shea, PhD, Royal Marsden NHS Foundation Trust, Radiotherapy Physics, London, UK

Helen A. McNair, PhD, Royal Marsden NHS Foundation Trust, Radiotherapy Department, London, UK

Sophie Alexander, MSc, Royal Marsden NHS Foundation Trust, Radiotherapy Department, London, UK

Eleftheria Kalaitzaki†, MSc, Royal Marsden NHS Foundation Trust, Clinical Trials Unit, London, UK

Jeffrey C. Bamber, PhD, Institute of Cancer Research, Radiotherapy and Imaging, London, UK

Emma J. Harris*, PhD, Institute of Cancer Research, Radiotherapy and Imaging, London, UK

Susan Lalondrelle*, FRCR, Royal Marsden NHS Foundation Trust, Radiotherapy Department, London, UK

*Joint last author

† Author responsible for statistical analyses

Corresponding author: Dr Emma Harris

Mailing Address: Department of Radiotherapy and Imaging, Institute of Cancer Research, 15 Cotswold Rd, Sutton, London, SM2 5NG

Phone number: +442086613320

Fax number: 02086433812

Email: emma.harris@icr.ac.uk

Acknowledgement of grant/financial support:

Research at The Institute of Cancer Research is supported by Cancer Research UK under Programme Grants C33389/A19727 and C20892/A23557. We acknowledge NHS funding to the NIHR Biomedical Research Centre at The Royal Marsden and The Institute of Cancer Research.

Conflict of interest: none

Full title: Combined ultrasound and cone beam CT improves target segmentation for image guided radiation therapy in uterine cervix cancer

Short running title: Target localization with CBCT-US fusion

Conflicts of interest notification:

None

ACCEPTED MANUSCRIPT

Abstract:

Purpose:

Adaptive radiation therapy strategies could account for interfractional uterine motion observed in cervix cancer patients, but the current CBCT-based treatment workflow is limited by poor soft tissue contrast. The goal of the present study was to determine if ultrasound (US) could be used to improve visualisation of the uterus, either as a single modality or in combination with CBCT.

Methods and materials:

Interobserver uterine contour agreement and confidence were compared on 40 corresponding CBCT, US, and CBCT-US-fused images from 11 cervix cancer patients. Contour agreement was measured using the Dice Similarity Coefficient (DSC) and mean contour-to-contour distance (MCCD). Observers rated their contour confidence on a scale from 1 to 10. Pairwise Wilcoxon signed-rank tests were used to measure differences in contour agreement and confidence.

Results:

CBCT-US-fused images had significantly better contour agreement and confidence than either individual modality ($p < 0.05$), with median [interquartile range (IQR)] values of 0.84 [0.11], 1.26 [0.23] mm, and 7 [2] for the DSC, MCCD, and observer confidence ratings, respectively. Contour agreement was similar between US and CBCT, with median [IQR] DSCs of 0.81 [0.17] and 0.82 [0.14], and MCCDs of 1.75 [1.15] mm and 1.62 [0.74] mm. Observers were significantly more confident in their US-based contours than their CBCT-based contours ($p <$

0.05), with median [IQR] confidence ratings of 7 [2.75] versus 5 [4].

Conclusions:

CBCT and US are complementary, and improve uterine segmentation precision when combined. Observers could localize the uterus with a similar precision on independent US and CBCT images.

Key words: ultrasound, adaptive radiotherapy, uterine cervix cancer, cone-beam CT, segmentation

Introduction

The primary clinical target volume (CTV_p) in cervix radiation therapy (RT) includes the uterus and cervix¹, which are highly mobile structures. Interfraction motion ranging from 2 mm to 60 mm has been observed due to changes in bladder and rectal volume and tumor regression²⁻⁵. Generous CTV_p to planning target volume (PTV) margins are necessary to compensate for positional uncertainty, but this increases the volume of normal tissue within the prescription dose region, which increases the risk of toxicity⁶⁻¹⁰.

Cone beam computed tomography (CBCT) is now widely available for bone, fiducial^{11,12} and soft tissue-based treatment verification¹³⁻¹⁹, and plan of the day (POD) selection^{19,20}. Disadvantages of CBCT include additional radiation dose and poor soft tissue contrast due to scatter and reconstruction artifacts. Indeed, image quality on 12% to 18% of gynaecological and prostate CBCT has been reported to be too poor for soft-tissue visualization purposes¹³⁻¹⁸. Even on CBCTs deemed suitable for soft-tissue analyses, poor image quality is reported to be the source of high interobserver contouring variability of the uterus and prostate^{15,17}.

Ultrasound (US) could be an effective, non-ionising, low-cost solution for providing high quality images of the pelvic anatomy at the time of radiation treatment delivery. US is routinely used in gynaecological/obstetric applications, and could be a promising alternative or adjunct to CBCT for image-guidance in cervix cancer radiation therapy^{11,21,22}. Baker et al. (2014) observed the apparent superiority of US image quality compared with CBCT in visualizing the uterus in a study investigating interfractional uterine motion, but no quantitative analysis

was performed¹¹. However, there are some drawbacks associated with US imaging of the uterus and cervix. (Note: for the remainder of this text, the uterus and cervix complex will be considered a single structure, and will be referred to as the uterus). US image quality is operator dependent and varies depending on the amount of abdominal fat, bladder volume, the amount of probe pressure applied, and whether there are obstructions such as gas or bone in the beam path²³⁻²⁶. Additionally, mechanically-swept 3D probes have a relatively small field of view, making it difficult to capture the entire uterus (particularly in cases where the disease is bulky), and impossible to capture all of the nearby organs at risk, particularly the rectum and bowel and involved lymph nodes within a single sweep.

A quantitative comparison of uterine segmentation on US and CBCT has not been evaluated previously. We expected inter-patient variability in uterus visualization and that either modality may be superior depending on patient characteristics at the time of treatment. It is not known if combining/fusing CBCT and US provides additional benefit in accuracy and precision of uterus localization for soft tissue-based treatment verification, POD selection, and on-line replanning.

The purpose of this work was to compare interobserver agreement in uterine segmentation and observer confidence in segmentation on CBCT, US, and CBCT-US-fused images to determine the optimal imaging method for target localization during cervix RT.

Methods and materials

Patients and treatment

Eleven patients with biopsy proven diagnosis of locally advanced cervix cancer were included in this NHS Research Ethics approved study (reference: 15/LO/1438). Fédération Internationale de Gynécologie Obstétrique (FIGO) stage distribution was: IIA= 1, IIB= 9, IIIB= 0, IVA= 1. The mean patient age was 51 (\pm 16) years. Patients were treated with radical chemoradiotherapy from February 2016 to May 2017. Patients were instructed to drink 350 ml water after complete bladder voiding 1 hour before treatment as per institutional protocol to maintain interfractional bladder volume and consistent set up. There were no bowel preparation instructions. Kilovoltage CBCT images were acquired immediately before treatment for on-line correction based on bony registration days 1-3 and weekly thereafter unless there was a systematic error of >5 mm or clinical indication, in which case they were acquired more frequently. 3D US of the uterus were acquired using the Clarity® system immediately before CBCT acquisition at 4-6 treatment sessions following patient set up on the RT treatment couch.

Data acquisition:

US imaging - 3D US (5 MHz center frequency, mechanically-swept probe) was acquired using the Clarity® system (Elekta Ltd, Stockholm, Sweden). The Clarity® system is described in detail elsewhere²⁷, but briefly, it is a standard US imaging system that is integrated into the RT clinic via infrared tracking, whereby the position of the probe (and the corresponding US images) with respect to the isocenter of the treatment room is known with sub-millimeter accuracy. US operators (either a clinical oncologist or a therapeutic

radiographer) applied a thick layer of US gel to the probe and scanned the uterus transabdominally using the smallest probe pressure possible to minimize soft-tissue deformation while still obtaining clear visualization of the uterus.

CBCT imaging - CBCT imaging (Elekta Ltd, Stockholm, Sweden) was performed immediately after US scanning, with no more than a five-minute interval between US and CBCT scans. CBCT imaging parameters were 120 kVp and 80 mAs with 350 projections and a bowtie filter.

64 US-CBCT image pairs (128 images in total) were obtained as part of this study. Two image pairs were excluded due to US operator errors in the probe calibration step of the Clarity QA, which caused misregistration of US images to the treatment room isocenter. Five image pairs were excluded due to failure to save US or CBCTs. Of the 57 remaining image pairs, 40 were randomly selected for analysis, leaving the remainder (17 image pairs) exclusively for observer training purposes.

Image formatting:

Image registration- CBCT images were registered to the planning CT scan using the Synergy bone match algorithm. Translational and rotational error was summarized as a translational couch shift. Following export to the Clarity workstation these translational shifts were applied to each CBCT to replicate match to planning CT. The infrared tracking technology provided by the Clarity system enabled the spatial registration between the US images and the CBCT images. In the offline Clarity workstation, the same translational moves were applied to the corresponding US images.

Image presentation - A software application was written in Matlab (Mathworks, Natick MA) to enable presentation of CT-CBCT, CT-US and CBCT-US registration for assessment by observers. The pixel size (in the format [superior-inferior direction, anterior-posterior direction]) was [2.5, 1] mm for CBCT images, [0.58, 0.58] mm for US, and [3, 1] mm for CT. Registered images were superimposed over one another, with user-adjustable sliders for 3D slice selection, transparency and windowing in both the sagittal and axial orientations. This functionality was achieved by interpolating the image in the registration with the larger pixels to a grid matching the sampling density of the image with the smaller pixels. Note that any references to 'CBCT', 'US' hereafter imply registration with the planning CT, as this is standard clinical practice. In the case of CBCT-US registration (henceforth referred to as CBCT-US-fusion), the planning CT was available for reference in a separate window rather than as a third superimposed image.

Image contouring and rating:

Observers used the MATLAB application for all image analysis including:

1. Delineation of the uterus on a pre-selected 2D sagittal slice: The sagittal slice used for contouring was the center-most slice of the uterus, as identified by one of the observers in the Experienced cohort (described in next section). This slice was in the same position for corresponding CBCT, US, and CBCT-US-fusion images.
2. Contour confidence ratings: After contouring the uterus, each observer was asked to rate the confidence that their contour reflected the true uterine

boundary from 1 (extremely unconfident) to 10 (extremely confident).

Observers first evaluated CBCT and US images separately. These images were displayed in a random order so that observers were blind to which CBCTs and US images were paired, and which images were from the same patient. Then, observers evaluated the CBCT-US-fusion images, which were also displayed in a random order.

One experienced observer also contoured the bladder on the central 2D sagittal slice on US and CBCT so that the relationship between bladder size and image quality could be assessed.

Observer selection and training:

Eight observers assessed US, CBCT, and CBCT-US-fusion image quality in this study. Observers were divided into two cohorts:

1. Experienced (three observers): One clinician (XX) and two medical physicists (XXX and XX) with previous experience in interpreting both US and CBCT images.
2. New-to-ultrasound (five observers): Clinicians, radiographers, and medical physicists who may have had experience in CBCT image analysis, but no prior training in interpreting US images.

Prior to performing the analysis, all eight observers participated in a one-hour training session (using the training dataset comprised of 17 CBCT, US, and CBCT-US-fusion images) designed to (1) teach observers how to use the MATLAB software application, (2) give observers experience in interpreting and contouring US, CBCT, and CBCT-US-fusion images of the uterus and (3) establish

a consensus for rating contour confidence.

Contour agreement metrics:

The two geometric measures used to assess contour agreement were the Dice similarity coefficient (DSC)²⁸ and the mean-contour-to-contour distance (MCCD). For two contours X and Y, the DSC was calculated as $(2|X \cap Y|)/(|X| + |Y|)$ with 0 and 1 representing no overlap and perfect overlap respectively, and the MCCD was defined as

$$MCCD = \frac{1}{n} \sum_{i=1}^n |X_i - Y_i| \quad (1)$$

where n is the number of points comprising contour X²⁹.

Generation of the gold standard contour:

For every image analyzed, a single contour was generated by combining the three contours from each of the observers in the ‘Experienced’ cohort using Simultaneous Truth and Performance Level Estimation (STAPLE)³⁰. The gold standard contour generated by the STAPLE method is the contour that optimizes the sensitivity and specificity of all of the contours from each individual observer. As 40 CBCTs, 40 US images, and 40 CBCT-US-fusion images were analyzed, this resulted in the generation of 120 gold standard contours. The contour agreement between the remaining five contours from the ‘New-to-ultrasound’ cohort and the gold standard STAPLE contour was measured for each of the 120 images.

Observer ratings:

Observers were required to rate how confident they were that their contours

reflected the true uterine boundary on a ten-point scale. Observers were asked to consider all factors potentially influencing their rating, including: the visibility of the uterine boundary, the clarity of the bladder wall and bowel gas, the regularity of the uterine shape, and the position of anatomical landmarks with respect to the PTV. All observers had practice in using these factors to inform their rating during the one-hour training session.

Statistical analyses:

The median and interquartile range (IQR) DSC and MCCD between each of the five uterine contours in the 'New-to-ultrasound' cohort and the gold standard STAPLE contour were reported for CBCT, US, and CBCT-US-fusion images. Pairwise Wilcoxon signed-rank tests³¹ with Bonferroni correction were performed to determine whether there were differences in the (1) DSC and (2) MCCD between imaging methods (US, CBCT, and CBCT-US-fusion). The statistical analysis described above was also repeated for comparing the observer contour confidence ratings from all eight observers from both the 'Experienced' and 'New-to-ultrasound' cohorts in CBCT, US, and CBCT-US-fused images.

The interobserver contour agreement and contour confidence was also evaluated in the Experienced cohort to assess the influence of observer training on the results. A pairwise analysis was used to calculate the median and IQR DSC, MCCD, and confidence ratings of the contours from these three observers. Differences between US, CBCT, and CBCT-US-fusion were detected using a pairwise Wilcoxon signed-rank test with Bonferroni correction.

Spearman's rank order correlation coefficient (R_s) and the corresponding p value were reported to assess the direction, strength, and significance of the association between the bladder area and the median contour confidence rating from all eight observers for independent CBCT and US images.

Results

The median [IQR] DSC and MCCD between observer contours from the New-to-US cohort and the gold standard STAPLE contours are shown in Table 1 for CBCT, US, and CBCT-US-fusion images. The DSC between observer contours on CBCT-US-fusion images was significantly greater than the DSC between observer contours on CBCT ($p = 0.002$) and US images ($p = 7.5e^{-4}$). The MCCD between observer contours on CBCT-US-fusion images was statistically significantly lower than CBCT images ($p = 3.4e^{-20}$) and US images ($p = 3.6e^{-24}$). The MCCD was statistically significantly lower on CBCT images than US images ($p = 6.5e^{-4}$).

Median [IQR] contour confidence ratings were 5 [4], 7 [2.75], and 7 [2] for CBCT, US, and CBCT-US-fusion images, respectively. The results from the Wilcoxon signed-rank tests revealed that (a) observers were significantly more confident in their contours drawn on CBCT-US-fusion images compared with CBCT and US (p values of $2.7e^{-28}$ and 0.005, respectively) and (b) observers were significantly more confident in their contours drawn on US than CBCT ($p = 4.5e^{-14}$).

The median [IQR] DSC, MCCD, and observer ratings from the pairwise analysis of the 'Experienced' cohort are shown in Table 2 for CBCT, US, and CBCT-US fusion images. There was no significant difference between the DSC on US and CBCT ($p = 0.97$), but the DSC between contours on the CBCT-US-fusion was significantly

higher than both CBCT and US (p values of $5.8e^{-4}$ and $4.2e^{-4}$ respectively). All MCCD values and confidence ratings were significantly different between experienced observers on CBCT, US, and CBCT-US-fusion images ($p < 5e^{-5}$), with the CBCT-US-fusion consistently outperforming CBCT and US.

Spearman's correlation coefficient (R_s) between bladder area and median contour confidence was -0.12 ($p = 0.83$) and 0.37 ($p = 0.02$) for CBCT and US images respectively.

Discussion

This study has demonstrated the superiority of combined US-CBCT imaging over either single modality for in room target localization of the uterus during radiation therapy for cervix cancer. We have also shown that US alone is comparable to CBCT as a sole modality.

CBCT versus US:

There was no statistically significant difference in DSC between the two modalities, CBCT had a statistically significantly smaller MCCD than US and observers were more confident in their US-based contours than their CBCT-based contours. This indicates that in general, CBCT and US have similar image qualities when used to visualize the uterus prior to RT in cervix cancer patients.

The reason that the higher observer ratings on US did not correspond to an improvement in interobserver contouring agreement in the new-to-ultrasound cohort is unclear, but could be due to the limited amount of observer experience in interpreting US (one hour) compared with CBCT (routine clinical use). This is

corroborated by the analyses performed on the Experienced cohort, where the improved observer confidence rating in ultrasound image quality did correspond to a significantly improved MCCD compared to CBCT. It is well known that there is a steep learning curve in US image interpretation partially due to the fact that observers can't rely on the body habitus or skeleton to orient themselves with the internal anatomy. Furthermore, inspection of the data also revealed that the observers in the 'New-to-ultrasound' cohort sometimes misinterpreted the non-homogeneous soft tissue contrast, uterine sub-structures (such as cysts or the endometrial lining) or US artifacts as the uterine boundary; errors which could potentially be overcome with more training. Examples of such US artifacts are shown side by side with the corresponding CBCT with and without observer contours in Figure 1. Even with these challenges, interobserver contouring agreement on US in the 'New-to-US' observer cohort matched that of CBCT after only a one-hour training module.

In this work, potential geometric mismatches between US and CBCT were not explicitly accounted for: factors such as probe pressure, gas movement, bulk patient motion, etc. could cause positional discrepancies between the apparent position of the uterus on the two modalities. However, preliminary measurements from this study did not detect a spatial mismatch: the median [IQR] DSC between the CBCT and US gold standard STAPLE contours was 0.78 [0.09], which is similar to the interobserver contouring variability of the uterus measured in previous work (median [IQR] DSC of 0.78 [0.11])³².

As the bladder is situated anteriorly to the uterus, its volume can influence US image quality: a full bladder can provide an acoustic window to the gynecologic

anatomy as urine has a low acoustic attenuation^{33,34}. Although the correlation coefficient between bladder area and contour confidence was relatively low (0.37), it was statistically significant and in the expected direction, indicating that observer contour confidence increases with increasing bladder size. Aside from bladder size, other factors such as body mass index (BMI), age, and the presence of gas in the acoustic path could also affect US image quality, which may explain why the relationship between bladder size and US image quality was only weakly detected in this study. In comparison, a lesser association was expected between bladder size and CBCT image quality, which was reflected by the low, statistically insignificant R_s value between bladder size and contour confidence ratings on CBCT of -0.12. The relationship between bladder size and median contour confidence ratings is displayed in Figure 2 for CBCT and US images.

In cases where the uterus is clearly visualized on US, US could potentially replace CBCT for plan verification and plan of the day adaptive treatment for cervix cancer patients. US could also be used for image-guidance in centers without CBCT imaging systems (i.e. in low to middle income countries), or to reduce concomitant imaging dose from frequent image-guidance.

CBCT-US-fusion:

According to all of the metrics considered in this study (DSC, MCCD, and observer ratings), CBCT-US-fusion images enabled observers to localize the uterus with more precision and confidence than either modality on its own. In some cases US provided better visualization of the uterus, and in others CBCT did, as shown in Table 3, which compares the median DSC and MCCD values for

each individual time point. Therefore, acquiring images from both modalities increases the probability of obtaining sufficient image quality to identify the uterus, which is one reason why the CBCT-US-fusion outperformed US and CBCT alone (see Figure 3 columns 1 - 2). However, the dominant benefit of combining CBCT and US images lies in the fact that they provide complementary information, and are most powerful when used together. This is clearly demonstrated by the MCCD, which was lowest in the CBCT-US-fusion images 75% of the time compared with either individual modality (see Table 3). The last column of Figure 4 exemplifies the complementary nature of CBCT and US, as only the superior-inferior extent of the uterus is visible on CBCT, whereas only the anterior-posterior extent of the uterus is visible on US: the CBCT-US-fusion enabled observers to combine the information from both modalities to determine the uterine boundary in all directions much more precisely. Columns 3 and 4 of Figure 4 further demonstrate the ability of observers to improve contouring precision on CBCT-US-fused images, even when the uterine boundary on both modalities individually is unclear.

Rationale for contouring in 2D rather than 3D:

It has been established that uterine motion occurs predominantly in the superior-inferior and anterior-posterior anatomical directions^{5,35}, and as such, this motion may be fully captured by 2D imaging along a central sagittal plane. As this plane is also the most informative part of the 3D dataset, both in terms of establishing uterine position and for its information content for defining the uterine boundary by US, it has been used to manually initialize automated 3D uterine segmentation in Clarity's® Assisted Gyne Segmentation (AGS)

algorithm^{32,36}. As such automated segmentation algorithms would need to be employed in an adaptive RT workflow to minimize contouring burden, 2D manual contour agreement on the central sagittal slice would be representative of uterine localization accuracy achievable in the clinic.

Feasibility of integrating US into the clinical workflow

Integrating US into the RT workflow was easily achieved using the Clarity® System, which enables the spatial registration of US images to treatment room coordinates via infrared probe tracking technology. This made fusion with CBCT and CT trivial, as all images were referenced to the same coordinate system. The small size of the system enabled us to leave it in the RT treatment room when not in use, making it easily accessible when needed. Daily quality assurance time is around 5 minutes. US acquisition and registration was easily incorporated into a standard treatment time slot (10 minutes).

As uterine localization can be achieved using US with little impact on hospital resources, the benefits of daily US imaging for cervical cancer patients may outweigh the cost of additional equipment. However, as US is not appropriate for determining the position of elective or involved nodal targets, we propose it either as an adjunct to other imaging modalities such as CBCT, or as a convenient and non-invasive soft-tissue verification method on days when CBCT imaging is not routinely performed.

Future directions:

Concern with regard to probe pressure effects and geometric accuracy of US images can be explored using CT or MRI as the gold standard imaging modality.

The development of an US training program for users and semi-automated segmentation could improve speed and accuracy of US-based verification. It also remains to be determined if patient and tumor characteristics pre-treatment can predict the best imaging modality to employ. Future work could include modifying the software program such that it records the time it takes for observers to identify the uterus on CBCT, US, and CBCT-US-fusion images to see whether fusion improves observer speed as well as contour agreement.

1.6 Conclusions

Target localization for cervix cancer radiation therapy is similar for CBCT and US. A weak association between bladder size and contour confidence on US images was observed, suggesting that larger bladder sizes may improve US image quality of the uterus. The combination of CBCT and US improves further the precision of target localization. Initial findings indicate that in cases where the uterus is clearly visualized on US, US can be used as an alternative to CBCT for uterine localization, with the benefit of reduction in radiation exposure and cost-effectiveness. Combined CBCT-US imaging could be used to implement adaptive planning strategies for cervix cancer radiation therapy.

References:

1. Lim K, Small W, Portelance L, et al. Consensus guidelines for delineation of clinical target volume for intensity-modulated pelvic radiotherapy for the definitive treatment of cervix cancer. *Int J Radiat Oncol Biol Phys*. 2011;79(2):348-355. doi:10.1016/j.ijrobp.2009.10.075
2. Bondar ML, Hoogeman MS, Mens JW, et al. Individualized nonadaptive and online-adaptive intensity-modulated radiotherapy treatment strategies for cervical cancer patients based on pretreatment acquired variable bladder filling computed tomography scans. *Int J Radiat Oncol Biol Phys*. 2012;83(5):1617-1623. doi:10.1016/j.ijrobp.2011.10.011
3. Chan P, Dinniwell R, Haider M a, et al. Inter- and intrafractional tumor and organ movement in patients with cervical cancer undergoing radiotherapy: a cinematic-MRI point-of-interest study. *Int J Radiat Oncol Biol Phys*. 2008;70(5):1507-1515. doi:10.1016/j.ijrobp.2007.08.055
4. Collen C, Engels B, Duchateau M, et al. Volumetric imaging by megavoltage computed tomography for assessment of internal organ motion during radiotherapy for cervical cancer. *Int J Radiat Oncol Biol Phys*. 2010;77(5):1590-1595. doi:10.1016/j.ijrobp.2009.10.021
5. Jadon R, Pembroke C a, Hanna CL, et al. A systematic review of organ motion and image-guided strategies in external beam radiotherapy for cervical cancer. *Clin Oncol (R Coll Radiol)*. 2014;26(4):185-196. doi:10.1016/j.clon.2013.11.031

6. Brixey C, Roeske J, Lujan A, Lumada S. Impact of intensity modulated radiotherapy on haematologic toxicity in women with gynecologic malignancies. *Int J Radiat Oncol Biol Phys.* 2002;54(5):1388.
7. Khan A, Jensen LG, Sun S, et al. Optimized planning target volume for intact cervical cancer. *Int J Radiat Oncol Biol Phys.* 2012;83(5):1500-1505.
doi:10.1016/j.ijrobp.2011.10.027
8. Mell LK, Kochanski JD, Roeske JC, et al. Dosimetric predictors of acute hematologic toxicity in cervical cancer patients treated with concurrent cisplatin and intensity-modulated pelvic radiotherapy. *Int J Radiat Oncol Biol Phys.* 2006;66(5):1356-1365. doi:10.1016/j.ijrobp.2006.03.018
9. Mundt AJ, Mell LK, Roeske JC. Preliminary analysis of chronic gastrointestinal toxicity in gynecology patients treated with intensity-modulated whole pelvic radiation therapy. *Int J Radiat Oncol.* 2003;56(5):1354-1360. doi:10.1016/S0360-3016(03)00325-0
10. Roeske JC, Lujan A, Rotmensch J, Waggoner SE, Yamada D, Mundt AJ. Intensity-modulated whole pelvic radiation therapy in patients with gynecologic malignancies. *Int J Radiat Oncol Biol Phys.* 2000;48(5):1613-1621. doi:10.1016/S0360-3016(00)00771-9
11. Baker M, Jensen JA, Behrens CF. Determining inter-fractional motion of the uterus using 3D ultrasound imaging during radiotherapy for cervical cancer. *Proc SPIE.* 2014;9040:90400Y-90400Y-10.
doi:10.1117/12.2043173

12. Heijkoop ST, Langerak TR, Quint S, et al. Clinical Implementation of an Online Adaptive Plan-of-the-Day Protocol for Nonrigid Motion Management in Locally Advanced Cervical Cancer IMRT. *Int J Radiat Oncol Biol Phys.* 2014;90(3):673-679. doi:10.1016/j.ijrobp.2014.06.046
13. Heijkoop ST, Langerak TR, Quint S, et al. Quantification of intra-fraction changes during radiotherapy of cervical cancer assessed with pre- and post-fraction Cone Beam CT scans. *Radiother Oncol.* 2015;117(3). doi:10.1016/j.radonc.2015.08.034
14. Langerak T, Heijkoop S, Quint S, Mens J, Heijmen B, Hoogeman M. Towards Automatic Plan Selection for Radiotherapy of Cervical Cancer by Fast Automatic Segmentation of Cone Beam CT Scans. *Med Image Comput Assist Interv 2014.* 2014;17:528-535.
15. Lutgendorf-Caucig C, Fotina I, Stock M, Potter R, Goldner G, Georg D. Feasibility of CBCT-based target and normal structure delineation in prostate cancer radiotherapy: Multi-observer and image multi-modality study. *Radiother Oncol.* 2011;98:154-161. doi:10.1016/j.radonc.2010.11.016
16. Maemoto H, Toita T, Ariga T, Heianna J, Yamashiro T, Murayama S. Predictive factors of uterine movement during definitive radiotherapy for cervical cancer. *J Radiat Res.* 2016;58(3):1-8. doi:10.1093/jrr/rrw101
17. Tyagi NE, Lewis JH, Yashar CM, et al. Daily online cone beam computed tomography to assess interfractional motion in patients with intact cervical cancer. *Int J Radiat Oncol Biol Phys.* 2011;80(1):273-280.

- doi:10.1016/j.ijrobp.2010.06.003
18. Wang X, Yu M, Wang J, et al. An assessment of interfractional bladder, rectum and vagina motion in postoperative cervical cancer based on daily cone-beam computed tomography. *Mol Clin Oncol*. 2016;4(2):271-277. doi:10.3892/mco.2015.704
 19. Heijkoop ST, Langerak TR, Quint S, et al. Clinical Implementation of an Online Adaptive Plan-of-the-Day Protocol for Nonrigid Motion Management in Locally Advanced Cervical Cancer IMRT. *Int J Radiat Oncol Biol Phys*. 2014;90(3):673-679. doi:10.1016/j.ijrobp.2014.06.046
 20. Sharfo AWM, Breedveld S, Voet PWJ, et al. Validation of Fully Automated VMAT Plan Generation for Library-Based Plan-of-the-Day Cervical Cancer Radiotherapy. *PLoS One*. 2016;11(12):1-13. doi:10.1371/journal.pone.0169202
 21. Dietz HP. Ultrasound imaging of the pelvic floor . Part I : two-dimensional aspects. 2004;(December 2003):80-92. doi:10.1002/uog.939
 22. Salomon LJ, Alfirevic Z, Berghella V, et al. Practice guidelines for performance of the routine mid-trimester fetal ultrasound scan. *Ultrasound Obstet Gynecol*. 2011;37(1):116-126. doi:10.1002/uog.8831
 23. FargierVoiron M, Presles B, Pommier P, et al. Impact of probe pressure variability on prostate localization for ultrasound-based image-guided radiotherapy. *Radiother Oncol*. 2014;111(1):132-137. doi:10.1016/j.radonc.2014.02.008

24. Hernandez-andrade E, Romero R, Ahn H, et al. Transabdominal evaluation of uterine cervical length during pregnancy fails to identify a substantial number of women with a short cervix. *J Matern Neonatal Med.* 2012;25(9):1682-1689. doi:10.3109/14767058.2012.657278
25. Kossoff G, Griffiths KA, Dixon CE. Is the quality of transvaginal images superior to transabdominal ones under matched conditions? *Ultrasound Obstet Gynecol.* 1991;1:29-35.
26. Paladini D. Sonography in obese and overweight pregnant women: clinical, medicolegal and technical issues. *Ultrasound Obs Gynecol.* 2009;33(6):720-729. doi:10.1002/uog.6393
27. Lachaine M, Falco T. Intrafractional prostate motion management with the Clarity Autoscan system. *Med Phys Int.* 2013;1(1):72-80. doi:10.1017/CB09781107415324.004
28. Dice LR. Measures of the Amount of Ecologic Association Between Species. *Ecol Soc Am.* 1945;26(3):297-302. doi:10.2307/1932409
29. Yan P, Xu S, Turkbey B, Kruecker J. Discrete Deformable Model Guided by Partial Active Shape Model for TRUS Image Segmentation. *IEEE Trans Biomed Eng.* 2010;57(5):1158-1166.
30. Warfield SK, Zou KH, Wells WM. Simultaneous truth and performance level estimation (STAPLE): An algorithm for the validation of image segmentation. *IEEE Trans Med Imaging.* 2004;23(7):903-921. doi:10.1109/TMI.2004.828354

31. F W. Individual comparisons by ranking methods. *Biometrics*. 1945;1:80-83.
32. XXXXX
33. Aguilera P a, Choi T, Durham B a. Ultrasound-guided suprapubic cystostomy catheter placement in the emergency department. *J Emerg Med*. 2004;26(3):319-321. doi:10.1016/j.jemermed.2003.11.016
34. Hasgall P, Gennaro F, Baumgartner C, et al. IT'IS Database for thermal and electromagnetic parameters of biological tissues. doi:10.13099/VIP21000-03-0
35. Taylor A, Powell MEB. An assessment of interfractional uterine and cervical motion: Implications for radiotherapy target volume definition in gynaecological cancer. *Radiother Oncol*. 2008;88(2):250-257. doi:10.1016/j.radonc.2008.04.016
36. Elekta Ltd. Software 4.0 User Guide. 2014;0400.

Figure legends:

Figure 1: Examples of ultrasound (US) imaging features (row 1) that caused contouring disagreement (row 2) in Patients 2, 8, 11, and 7 (p2 – p11). Such features include non-homogenous soft-tissue contrast in the uterus, internal structures such as cysts and the endometrial lining, and US artefacts such as shadowing and reverberations. Corresponding CBCTs and CBCT-based contours shown for reference in rows 3 and 4. The US and CBCT images are displayed on different scales (pixel sizes of [0.58, 0.58] mm and [2.5, 1] mm, respectively).

Figure 2: Scatter plots demonstrating the relationship between bladder size (measured as the area of the bladder on a central sagittal slice) and the median observer contour confidence rating for CBCT images (top) and US images (bottom).

Figure 3: Comparison of interobserver contour agreement on CBCT (pixel size [1, 1, 1] mm), ultrasound (US) (pixel size [0.58, 0.58] mm), and CBCT-US-fusion images (pixel size [0.58, 0.58] mm). Each observer's contour is denoted by the colour given in the legend shown in the bottom row (Experienced observers = 1 - 3, 'New-to-US' observers = 4 - 8). Columns 1 - 2: examples where the benefit of CBCT-US-fusion was mainly due to the increased probability of acquiring an excellent image from one modality (denoted by the green check marks). Columns 3 - 5: examples where CBCT and US images provided complementary information, which is why the interobserver agreement is relatively poor on both modalities individually, but good on the CBCT-US-fusion image.

Table 1: Interobserver contour agreement and observer contour confidence results

Median and interquartile range [IQR] results for: (*columns 1 and 2*) the Dice similarity coefficient (DSC) and mean contour-to-contour distance (MCCD) between the gold standard contour and the 5 contours from the new-to-US cohort and (*column 3*) observer ratings of contour confidence (1 = extremely unconfident and 10 = extremely confident) from all 8 observers. Symbols indicate significant differences ($p < 0.005$) between imaging modalities within a column (Ψ indicates a significant difference between CBCT and US and ** indicates that CBCT-US-fusion is significantly different to both CBCT and US).

	DSC	MCCD (mm)	Observer rating
CBCT	0.82 [0.14]	1.63 [0.74] Ψ	5 [4]
Ultrasound (US)	0.81 [0.17]	1.75 [1.15]	7 [2.75] Ψ
CBCT-US-fusion	0.84 [0.11]**	1.26 [0.23]**	7 [2]**

Table 2: Pairwise interobserver contour agreement and confidence from Experienced observer cohort (3 observers).


Median and interquartile range [IQR] results from the pairwise analysis from the 3 observers in the Experienced cohort for: (*column 1*) the Dice similarity coefficient (DSC), (*column 2*) the mean contour-to-contour distance (MCCD) and (*column 3*) observer ratings of contour confidence (1 = extremely unconfident and 10 = extremely confident). Symbols indicate significant differences ($p < 0.005$) between imaging modalities within a column (Ψ indicates a significant difference between CBCT and US and ** indicates that CBCT-US-fusion is significantly different to both CBCT and US).



	DSC	MCCD (mm)	Observer rating
CBCT	0.82 [0.13]	2.94 [1.58]	4 [4]
Ultrasound (US)	0.81 [0.14]	2.25 [2.67] Ψ	7 [3] Ψ
CBCT-US-fusion	0.87 [0.07]**	1.63 [0.63]**	8 [2]**

Table 3: Highest level contour agreement categorized by imaging method

Frequency that each modality (CBCT, ultrasound (US) and CBCT-US-fusion) had the highest median DSC (*column 1*) and lowest median MCCD (*column 2*) in absolute cases and as a percentage of total time points (n = 40).




	DSC	MCCD
CBCT	11 (27.5%)	5 (12.5%)
Ultrasound (US)	13 (32.5%)	5 (12.5%)
CBCT-US-fusion	16 (40%)	30 (75%)










ACCEPTED






CBCCT only



US only



CBCCT-US fusion



Summary:

Ultrasound (US) and CBCT were assessed (1) as independent modalities and (2) in combination for localising the uterus in eleven cervix cancer patients for adaptive radiotherapy purposes. Interobserver uterine contour agreement was similar on independent US and CBCT images, though observers were significantly more confident in their US-based contours. However, both interobserver contour agreement and contour confidence were significantly improved when US and CBCT were combined, as the two modalities were shown to be complementary.